

Appendix E

## Calibration Certificates of Monitoring Equipments



專業化驗有限公司  
QUALITY PRO TEST-CONSULT LIMITED

Unit 10, 14/F, Wah Wai Centre, 38-40 Au Pui Wan St., Fotan, Hong Kong  
Email: info@qualityprotest.com; Website: www.qualityprotest.com  
Tel: (852) 3956 8717; Fax: (852) 3956 3928

## REPORT OF EQUIPMENT PERFORMANCE CHECK/ CALIBRATION

Report No. : AI090154  
Date of Issue : 02 October 2019  
Page No. : 1 of 2

### PART A – CUSTOMER INFORMATION

Enovative Environmental Service Ltd.  
Flat 2207, Yu Fun House,  
Yu Chui Court, Shatin  
New Territories, Hong Kong  
Attn: Mr. Thomas WONG

### PART B – DESCRIPTION

Name of Equipment : YSI ProDSS (Multi-Parameters)  
Manufacturer : YSI (a xylem brand)  
Serial Number : 16H104233  
Date of Received : Sep 27, 2019  
Date of Calibration : Sep 27, 2019  
Date of Next Calibration<sup>(a)</sup> : Dec 26, 2019

### PART C – REFERENCE METHODS/ DOCUMENTS FOR THE CALIBRATION

| Parameter            | Reference Method   |
|----------------------|--|
| pH at 25°C           | APHA 21e 4500-H <sup>+</sup> B   |
| Dissolved Oxygen     | APHA 21e 4500-O G  |
| Conductivity at 25°C | APHA 21e 2510 B  |
| Salinity             | APHA 21e 2520 B  |
| Turbidity            | APHA 21e 2130 B  |
| Temperature          | Section 6 of international Accreditation New Zealand Technical Guide no. 3 Second edition March 2008: Working Thermometer Calibration Procedure. |

### PART D – CALIBRATION RESULTS<sup>(b,c)</sup>

#### (1) pH at 25°C

| Target (pH unit) | Displayed Reading <sup>(d)</sup> (pH Unit) | Tolerance <sup>(e)</sup> (pH Unit) | Results      |
|------------------|--|------------------------------------|--------------|
| 4.00             | 4.03                                       | 0.03                               | Satisfactory |
| 7.42             | 7.44                                       | 0.02                               | Satisfactory |
| 10.01            | 10.06                                      | 0.05                               | Satisfactory |

Tolerance of pH should be less than  $\pm 0.20$  (pH unit)

#### (2) Temperature

| Reading of Ref. thermometer (°C) | Displayed Reading (°C) | Tolerance (°C) | Results      |
|----------------------------------|------------------------|----------------|--------------|
| 10.0                             | 10.0                   | 0.0            | Satisfactory |
| 22.0                             | 22.1                   | 0.1            | Satisfactory |
| 42.0                             | 42.2                   | 0.2            | Satisfactory |

Tolerance limit of temperature should be less than  $\pm 2.0$  (°C)

~ CONTINUED ON NEXT PAGE ~

#### Remark(s): -

- <sup>(a)</sup> The "Date of Next Calibration" is recommended according to best practice principals as practiced by QPT or quoted from relevant international standards.  
<sup>(b)</sup> The results relate only to the calibrated equipment as received  
<sup>(c)</sup> The performance of the equipment stated in this report is checked with independent reference material and results compared against a calibrated secondary source.  
<sup>(d)</sup> "Displayed Reading" denotes the figure shown on item under calibration/ checking regardless of equipment precision or significant figures.  
<sup>(e)</sup> The "Tolerance Limit" mentioned is referenced to YSI product specifications.

  
LEE Chun-ning, Desmond  
Senior Chemist



專業化驗有限公司

QUALITY PRO TEST-CONSULT LIMITED

Unit 10, 14/F, Wah Wai Centre, 38-40 Au Pui Wan St., Fotan, Hong Kong

Email: info@qualityprotest.com; Website: www.qualityprotest.com

Tel: (852) 3956 8717; Fax: (852) 3956 3928

## REPORT OF EQUIPMENT PERFORMANCE CHECK/ CALIBRATION

Report No. : AI090154  
Date of Issue : 02 October 2019  
Page No. : 2 of 2

### PART D – CALIBRATION RESULTS (Cont'd)

#### (3) Dissolved Oxygen

| Expected Reading (mg/L) | Displayed Reading (mg/L) | Tolerance (mg/L) | Results      |
|-------------------------|--------------------------|------------------|--------------|
| 0.78                    | 0.75                     | -0.03            | Satisfactory |
| 3.69                    | 3.98                     | 0.29             | Satisfactory |
| 5.77                    | 5.4                      | -0.37            | Satisfactory |
| 7.68                    | 7.82                     | 0.14             | Satisfactory |

Tolerance limit of dissolved oxygen should be less than  $\pm 0.50$  (mg/L)

#### (4) Conductivity at 25°C

| Conc. of KCl (M) | Expected Reading ( $\mu\text{S}/\text{cm}$ ) | Displayed Reading ( $\mu\text{S}/\text{cm}$ ) | Tolerance (%) | Results      |
|------------------|--|---|---------------|--------------|
| 0.001            | 146.9  | 138.9   | -5.45         | Satisfactory |
| 0.01             | 1412   | 1380  | -2.27         | Satisfactory |
| 0.1              | 12890  | 12834   | -0.43         | Satisfactory |
| 0.5              | 58670  | 57663   | -1.72         | Satisfactory |
| 1.0              | 111900                                       | 109858  | -1.82         | Satisfactory |

Tolerance limit of conductivity should be less than  $\pm 10.0$  (%)

#### (5) Salinity

| Expected Reading (g/L) | Displayed Reading (g/L) | Tolerance (%) | Results      |
|------------------------|-------------------------|---------------|--------------|
| 10                     | 10.16                   | 1.60          | Satisfactory |
| 20                     | 20.38                   | 1.90          | Satisfactory |
| 30                     | 30.47                   | 1.57          | Satisfactory |

Tolerance limit of salinity should be less than  $\pm 10.0$  (%)

#### (6) Turbidity

| Expected Reading (NTU) | Displayed Reading <sup>(f)</sup> (NTU) | Tolerance <sup>(g)</sup> (%) | Results      |
|------------------------|--|------------------------------|--------------|
| 0                      | 0.11                                   | --                           | Satisfactory |
| 10                     | 9.89                                   | -1.1                         | Satisfactory |
| 20                     | 19.82                                  | -0.9                         | Satisfactory |
| 100                    | 97.25                                  | -2.8                         | Satisfactory |
| 800                    | 780.16                                 | -2.5                         | Satisfactory |

Tolerance limit of turbidity should be less than  $\pm 10.0$  (%)

~ END OF REPORT ~

#### Remark(s): -

<sup>(f)</sup> "Displayed Reading" presents the figures shown on item under calibration/ checking regardless of equipment precision or significant figures.

<sup>(g)</sup> The "Tolerance Limit" mentioned is the acceptance criteria applicable for similar equipment used by Quality Pro Test-Consult Ltd. or quoted from relevant international standards.



專業化驗有限公司  
QUALITY PRO TEST-CONSULT LIMITED

Unit 10, 14/F, Wah Wai Centre, 38-40 Au Pui Wan St., Fotan, Hong Kong  
Email: info@qualityprotest.com; Website: www.qualityprotest.com  
Tel: (852) 3956 8717; Fax: (852) 3956 3928

## REPORT OF EQUIPMENT PERFORMANCE CHECK/ CALIBRATION

Report No. : AI090155  
Date of Issue : 02 October 2019  
Page No. : 1 of 2

### PART A – CUSTOMER INFORMATION

Enovative Environmental Service Ltd.  
Flat 2207, Yu Fun House,  
Yu Chui Court, Shatin  
New Territories, Hong Kong  
Attn: Mr. Thomas WONG

### PART B – DESCRIPTION

Name of Equipment : YSI ProDSS (Multi-Parameters)  
Manufacturer : YSI (a xylem brand)  
Serial Number : 17H105557  
Date of Received : Sep 27, 2019  
Date of Calibration : Sep 27, 2019  
Date of Next Calibration<sup>(a)</sup> : Dec 26, 2019

### PART C – REFERENCE METHODS/ DOCUMENTS FOR THE CALIBRATION

| Parameter            | Reference Method  |
|----------------------|---|
| pH at 25°C           | APHA 21e 4500-H <sup>+</sup> B  |
| Dissolved Oxygen     | APHA 21e 4500-O G   |
| Conductivity at 25°C | APHA 21e 2510 B   |
| Salinity             | APHA 21e 2520 B   |
| Turbidity            | APHA 21e 2130 B   |
| Temperature          | Section 6 of international Accreditation New Zealand Technical<br>Guide no. 3 Second edition March 2008: Working Thermometer Calibration Procedure. |

### PART D – CALIBRATION RESULTS<sup>(b,c)</sup>

#### (1) pH at 25°C

| Target (pH unit) | Displayed Reading <sup>(d)</sup> (pH Unit) | Tolerance <sup>(e)</sup> (pH Unit) | Results      |
|------------------|--|------------------------------------|--------------|
| 4.00             | 4.05                                       | 0.05                               | Satisfactory |
| 7.42             | 7.41                                       | -0.01                              | Satisfactory |
| 10.01            | 10.11                                      | 0.10                               | Satisfactory |

Tolerance of pH should be less than  $\pm 0.20$  (pH unit)

#### (2) Temperature

| Reading of Ref. thermometer (°C) | Displayed Reading (°C) | Tolerance (°C) | Results      |
|----------------------------------|------------------------|----------------|--------------|
| 10.0                             | 10.0                   | 0.0            | Satisfactory |
| 22.0                             | 22.1                   | 0.1            | Satisfactory |
| 42.0                             | 42.1                   | 0.1            | Satisfactory |

Tolerance limit of temperature should be less than  $\pm 2.0$  (°C)

~ CONTINUED ON NEXT PAGE ~

#### Remark(s): -

- <sup>(a)</sup> The "Date of Next Calibration" is recommended according to best practice principals as practiced by QPT or quoted form relevant international standards.  
<sup>(b)</sup> The results relate only to the calibrated equipment as received  
<sup>(c)</sup> The performance of the equipment stated in this report is checked with independent reference material and results compared against a calibrated secondary source.  
<sup>(d)</sup> "Displayed Reading" denotes the figure shown on item under calibration/ checking regardless of equipment precision or significant figures.  
<sup>(e)</sup> The "Tolerance Limit" mentioned is referenced to YSI product specifications.

  
LEE Chun-ning, Desmond  
Senior Chemist





## REPORT OF EQUIPMENT PERFORMANCE CHECK/ CALIBRATION

Report No. : AI090155  
Date of Issue : 02 October 2019  
Page No. : 2 of 2

### PART D – CALIBRATION RESULTS (Cont'd)

#### (3) Dissolved Oxygen

| Expected Reading (mg/L) | Displayed Reading (mg/L) | Tolerance (mg/L) | Results      |
|-------------------------|--------------------------|------------------|--------------|
| 0.78                    | 0.72                     | -0.06            | Satisfactory |
| 3.69                    | 4.01                     | 0.32             | Satisfactory |
| 5.77                    | 5.38                     | -0.39            | Satisfactory |
| 7.68                    | 7.80                     | 0.12             | Satisfactory |

Tolerance limit of dissolved oxygen should be less than  $\pm 0.50$  (mg/L)

#### (4) Conductivity at 25°C

| Conc. of KCl (M) | Expected Reading ( $\mu\text{S/cm}$ ) | Displayed Reading ( $\mu\text{S/cm}$ ) | Tolerance (%) | Results      |
|------------------|---------------------------------------|--|---------------|--------------|
| 0.001            | 146.9                                 | 138.2                                  | -5.92         | Satisfactory |
| 0.01             | 1412                                  | 1394                                   | -1.27         | Satisfactory |
| 0.1              | 12890                                 | 12855                                  | -0.27         | Satisfactory |
| 0.5              | 58670                                 | 57582                                  | -1.85         | Satisfactory |
| 1.0              | 111900                                | 109780                                 | -1.89         | Satisfactory |

Tolerance limit of conductivity should be less than  $\pm 10.0$  (%)

#### (5) Salinity

| Expected Reading (g/L) | Displayed Reading (g/L) | Tolerance (%) | Results      |
|------------------------|-------------------------|---------------|--------------|
| 10                     | 10.08                   | 0.80          | Satisfactory |
| 20                     | 20.41                   | 2.05          | Satisfactory |
| 30                     | 30.52                   | 1.73          | Satisfactory |

Tolerance limit of salinity should be less than  $\pm 10.0$  (%)

#### (6) Turbidity

| Expected Reading (NTU) | Displayed Reading <sup>(f)</sup> (NTU) | Tolerance <sup>(g)</sup> (%) | Results      |
|------------------------|--|------------------------------|--------------|
| 0                      | 0.10                                   | --                           | Satisfactory |
| 10                     | 9.94                                   | -0.6                         | Satisfactory |
| 20                     | 19.86                                  | -0.7                         | Satisfactory |
| 100                    | 97.43                                  | -2.6                         | Satisfactory |
| 800                    | 779.37                                 | -2.6                         | Satisfactory |

Tolerance limit of turbidity should be less than  $\pm 10.0$  (%)

~ END OF REPORT ~

Remark(s): -

<sup>(f)</sup> "Displayed Reading" presents the figures shown on item under calibration/ checking regardless of equipment precision or significant figures.

<sup>(g)</sup> The "Tolerance Limit" mentioned is the acceptance criteria applicable for similar equipment used by Quality Pro Test-Consult Ltd. or quoted from relevant international standards.



專業化驗有限公司  
QUALITY PRO TEST-CONSULT LIMITED

Unit 10, 14/F, Wah Wai Centre, 38-40 Au Pui Wan St., Fotan, Hong Kong  
Email: info@qualityprotest.com; Website: www.qualityprotest.com  
Tel: (852) 3956 8717; Fax: (852) 3956 3928

## REPORT OF EQUIPMENT PERFORMANCE CHECK/ CALIBRATION

Report No. : AI100183  
Date of Issue : 30 October 2019  
Page No. : 1 of 2

### PART A – CUSTOMER INFORMATION

Enovative Environmental Service Ltd.  
Flat 2207, Yu Fun House,  
Yu Chui Court, Shatin  
New Territories, Hong Kong  
Attn: Mr. Thomas WONG

### PART B – DESCRIPTION

Name of Equipment : YSI 6920V2 (Multi-Parameters)  
Manufacturer : YSI (a xylem brand)  
Serial Number : 00019CB2  
Date of Received : Oct 28, 2019  
Date of Calibration : Oct 28, 2019  
Date of Next Calibration<sup>(a)</sup> : Jan 27, 2020

### PART C – REFERENCE METHODS/ DOCUMENTS FOR THE CALIBRATION

| <u>Parameter</u>     | <u>Reference Method</u>   |
|----------------------|---|
| pH at 25°C           | APHA 21e 4500-H <sup>+</sup> B  |
| Dissolved Oxygen     | APHA 21e 4500-O G   |
| Conductivity at 25°C | APHA 21e 2510 B   |
| Salinity             | APHA 21e 2520 B   |
| Turbidity            | APHA 21e 2130 B   |
| Temperature          | Section 6 of international Accreditation New Zealand Technical<br>Guide no. 3 Second edition March 2008: Working Thermometer Calibration Procedure. |

### PART D – CALIBRATION RESULTS<sup>(b,c)</sup>

#### (1) pH at 25°C

| Target (pH unit) | Displayed Reading <sup>(d)</sup> (pH Unit) | Tolerance <sup>(e)</sup> (pH Unit) | Results      |
|------------------|--|------------------------------------|--------------|
| 4.00             | 3.95                                       | -0.05                              | Satisfactory |
| 7.42             | 7.36                                       | -0.06                              | Satisfactory |
| 10.01            | 9.93                                       | -0.08                              | Satisfactory |

Tolerance of pH should be less than  $\pm 0.20$  (pH unit)

#### (2) Temperature


| Reading of Ref. thermometer (°C) | Displayed Reading (°C) | Tolerance (°C) | Results      |
|----------------------------------|------------------------|----------------|--------------|
| 15.0                             | 15.1                   | 0.1            | Satisfactory |
| 25.0                             | 24.9                   | -0.1           | Satisfactory |
| 35.0                             | 34.9                   | -0.1           | Satisfactory |

Tolerance limit of temperature should be less than  $\pm 2.0$  (°C)

~ CONTINUED ON NEXT PAGE ~

#### Remark(s): -

- <sup>(a)</sup> The "Date of Next Calibration" is recommended according to best practice principals as practiced by QPT or quoted from relevant international standards.  
<sup>(b)</sup> The results relate only to the calibrated equipment as received  
<sup>(c)</sup> The performance of the equipment stated in this report is checked with independent reference material and results compared against a calibrated secondary source.  
<sup>(d)</sup> "Displayed Reading" denotes the figure shown on item under calibration/ checking regardless of equipment precision or significant figures.  
<sup>(e)</sup> The "Tolerance Limit" mentioned is referenced to YSI product specifications.

  
LEE Chun-ning, Desmond  
Senior Chemist



## REPORT OF EQUIPMENT PERFORMANCE CHECK/ CALIBRATION

Report No. : AI100183  
Date of Issue : 30 October 2019  
Page No. : 2 of 2

### PART D – CALIBRATION RESULTS (Cont'd)

#### (3) Dissolved Oxygen

| Expected Reading (mg/L) | Displayed Reading (mg/L) | Tolerance (mg/L) | Results      |
|-------------------------|--------------------------|------------------|--------------|
| 1.04                    | 0.80                     | -0.24            | Satisfactory |
| 4.10                    | 4.34                     | 0.24             | Satisfactory |
| 5.92                    | 5.94                     | 0.02             | Satisfactory |
| 7.81                    | 8.07                     | 0.26             | Satisfactory |

Tolerance limit of dissolved oxygen should be less than  $\pm 0.50$  (mg/L)

#### (4) Conductivity at 25°C

| Conc. of KCl (M) | Expected Reading ( $\mu\text{S}/\text{cm}$ ) | Displayed Reading ( $\mu\text{S}/\text{cm}$ ) | Tolerance (%) | Results      |
|------------------|--|---|---------------|--------------|
| 0.001            | 146.9  | 140.0   | -4.70         | Satisfactory |
| 0.01             | 1412   | 1394  | -1.27         | Satisfactory |
| 0.1              | 12890  | 12780   | -0.85         | Satisfactory |
| 0.5              | 58670  | 57927   | -1.27         | Satisfactory |
| 1.0              | 111900                                       | 110880  | -0.91         | Satisfactory |

Tolerance limit of conductivity should be less than  $\pm 10.0$  (%)

#### (5) Salinity

| Expected Reading (g/L) | Displayed Reading (g/L) | Tolerance (%) | Results      |
|------------------------|-------------------------|---------------|--------------|
| 10                     | 9.90                    | -1.00         | Satisfactory |
| 20                     | 19.88                   | -0.60         | Satisfactory |
| 30                     | 29.89                   | -0.37         | Satisfactory |

Tolerance limit of salinity should be less than  $\pm 10.0$  (%)

#### (6) Turbidity

| Expected Reading (NTU) | Displayed Reading <sup>(f)</sup> (NTU) | Tolerance <sup>(g)</sup> (%) | Results      |
|------------------------|--|------------------------------|--------------|
| 0                      | 0.20                                   | --                           | Satisfactory |
| 10                     | 9.98                                   | -0.2                         | Satisfactory |
| 20                     | 19.88                                  | -0.6                         | Satisfactory |
| 100                    | 100.20                                 | 0.2                          | Satisfactory |
| 800                    | 798.82                                 | -0.1                         | Satisfactory |

Tolerance limit of turbidity should be less than  $\pm 10.0$  (%)

~ END OF REPORT ~

**Remark(s): -**

<sup>(f)</sup> "Displayed Reading" presents the figures shown on item under calibration/ checking regardless of equipment precision or significant figures.

<sup>(g)</sup> The "Tolerance Limit" mentioned is the acceptance criteria applicable for similar equipment used by Quality Pro Test-Consult Ltd. or quoted from relevant international standards.





專業化驗有限公司  
QUALITY PRO TEST-CONSULT LIMITED

Unit 10, 14/F, Wah Wai Centre, 38-40 Au Pui Wan St., Fotan, Hong Kong  
Email: info@qualityprotest.com; Website: www.qualityprotest.com  
Tel: (852) 3956 8717; Fax: (852) 3956 3928

## REPORT OF EQUIPMENT PERFORMANCE CHECK/ CALIBRATION

Report No. : AI100182  
Date of Issue : 30 October 2019  
Page No. : 1 of 2

### PART A – CUSTOMER INFORMATION

Enovative Environmental Service Ltd.  
Flat 2207, Yu Fun House,  
Yu Chui Court, Shatin  
New Territories, Hong Kong  
Attn: Mr. Thomas WONG

### PART B – DESCRIPTION

Name of Equipment : YSI 6920V2 (Multi-Parameters)  
Manufacturer : YSI (a xylem brand)  
Serial Number : 0001C6A7  
Date of Received : Oct 28, 2019  
Date of Calibration : Oct 28, 2019  
Date of Next Calibration<sup>(a)</sup> : Jan 27, 2020

### PART C – REFERENCE METHODS/ DOCUMENTS FOR THE CALIBRATION

| <u>Parameter</u>     | <u>Reference Method</u>  |
|----------------------|--|
| pH at 25°C           | APHA 21e 4500-H <sup>+</sup> B   |
| Dissolved Oxygen     | APHA 21e 4500-O G  |
| Conductivity at 25°C | APHA 21e 2510 B  |
| Salinity             | APHA 21e 2520 B  |
| Turbidity            | APHA 21e 2130 B  |
| Temperature          | Section 6 of international Accreditation New Zealand Technical Guide no. 3 Second edition March 2008: Working Thermometer Calibration Procedure. |

### PART D – CALIBRATION RESULTS<sup>(b,c)</sup>

#### (1) pH at 25°C

| Target (pH unit) | Displayed Reading <sup>(d)</sup> (pH Unit) | Tolerance <sup>(e)</sup> (pH Unit) | Results      |
|------------------|--|------------------------------------|--------------|
| 4.00             | 4.07                                       | 0.07                               | Satisfactory |
| 7.42             | 7.49                                       | 0.07                               | Satisfactory |
| 10.01            | 10.05                                      | 0.04                               | Satisfactory |

Tolerance of pH should be less than  $\pm 0.20$  (pH unit)

#### (2) Temperature


| Reading of Ref. thermometer (°C) | Displayed Reading (°C) | Tolerance (°C) | Results      |
|----------------------------------|------------------------|----------------|--------------|
| 15.0                             | 15.1                   | 0.1            | Satisfactory |
| 25.0                             | 25.0                   | 0.0            | Satisfactory |
| 35.0                             | 35.0                   | 0.0            | Satisfactory |

Tolerance limit of temperature should be less than  $\pm 2.0$  (°C)

~ CONTINUED ON NEXT PAGE ~

#### Remark(s): -

- <sup>(a)</sup> The "Date of Next Calibration" is recommended according to best practice principals as practiced by QPT or quoted from relevant international standards.  
<sup>(b)</sup> The results relate only to the calibrated equipment as received  
<sup>(c)</sup> The performance of the equipment stated in this report is checked with independent reference material and results compared against a calibrated secondary source.  
<sup>(d)</sup> "Displayed Reading" denotes the figure shown on item under calibration/ checking regardless of equipment precision or significant figures.  
<sup>(e)</sup> The "Tolerance Limit" mentioned is referenced to YSI product specifications.

  
LEE Chun-ning, Desmond  
Senior Chemist





## REPORT OF EQUIPMENT PERFORMANCE CHECK/ CALIBRATION

Report No. : AI100182  
Date of Issue : 30 October 2019  
Page No. : 2 of 2

### PART D – CALIBRATION RESULTS (Cont'd)

#### (3) Dissolved Oxygen

| Expected Reading (mg/L) | Displayed Reading (mg/L) | Tolerance (mg/L) | Results      |
|-------------------------|--------------------------|------------------|--------------|
| 1.04                    | 0.90                     | -0.14            | Satisfactory |
| 4.10                    | 4.40                     | 0.3              | Satisfactory |
| 5.92                    | 6.00                     | 0.08             | Satisfactory |
| 7.81                    | 8.10                     | 0.29             | Satisfactory |

Tolerance limit of dissolved oxygen should be less than  $\pm 0.50$  (mg/L)

#### (4) Conductivity at 25°C

| Conc. of KCl (M) | Expected Reading ( $\mu\text{S}/\text{cm}$ ) | Displayed Reading ( $\mu\text{S}/\text{cm}$ ) | Tolerance (%) | Results      |
|------------------|--|---|---------------|--------------|
| 0.001            | 146.9  | 156   | 6.19          | Satisfactory |
| 0.01             | 1412   | 1384  | -1.98         | Satisfactory |
| 0.1              | 12890  | 12810   | -0.62         | Satisfactory |
| 0.5              | 58670  | 57991   | -1.16         | Satisfactory |
| 1.0              | 111900                                       | 110844  | -0.94         | Satisfactory |

Tolerance limit of conductivity should be less than  $\pm 10.0$  (%)

#### (5) Salinity

| Expected Reading (g/L) | Displayed Reading (g/L) | Tolerance (%) | Results      |
|------------------------|-------------------------|---------------|--------------|
| 10                     | 10.08                   | 0.80          | Satisfactory |
| 20                     | 20.07                   | 0.35          | Satisfactory |
| 30                     | 30.1                    | 0.33          | Satisfactory |

Tolerance limit of salinity should be less than  $\pm 10.0$  (%)

#### (6) Turbidity

| Expected Reading (NTU) | Displayed Reading <sup>(f)</sup> (NTU) | Tolerance <sup>(g)</sup> (%) | Results      |
|------------------------|--|------------------------------|--------------|
| 0                      | 0.50                                   | --                           | Satisfactory |
| 10                     | 10.02                                  | 0.2                          | Satisfactory |
| 20                     | 20.47                                  | 2.3                          | Satisfactory |
| 100                    | 100.16                                 | 0.2                          | Satisfactory |
| 800                    | 798.93                                 | -0.1                         | Satisfactory |

Tolerance limit of turbidity should be less than  $\pm 10.0$  (%)

~ END OF REPORT ~

#### Remark(s): -

<sup>(f)</sup> "Displayed Reading" presents the figures shown on item under calibration/ checking regardless of equipment precision or significant figures.

<sup>(g)</sup> The "Tolerance Limit" mentioned is the acceptance criteria applicable for similar equipment used by Quality Pro Test-Consult Ltd. or quoted from relevant international standards.