

Appendix G

Calibration Certificate of Monitoring Equipment

Calibration Certificate

GD 005

Cert. Ref. No.: BW/XT/B/12943

Date: 2018 04 10

Customer: Gammon Construction Ltd.

Purchase Order No.: 1578603 OL

28/F., Devon House,

Date: 2018 04 12 INVOICE NO: AP

Taikoo Place 979 King's Road,

Email: Jessica.Lau@gammonconstruction.com

Hong Kong.

Attn: Jessica Lau

Tel: 3520 0392

Fax: 3520 0488

Mobile Phone

User Details:

Gas Detector Model: XT-XWHM-Y-OR

Serial No.: MA218-000709

Pump S/N: 101801

Calibration Record:

Inspection before calibration	Visual inspection	Functional Test
Basic Unit - Case, Clip & Display etc.	OK	OK
Battery and charge etc.	OK	OK
Authorized Pump	OK	OK
Other items		

Type of Sensor	Expiry Date		
Oxygen Sensor	2020	04	12
CO & H2S Sensor	2020	04	12
Combustible(LEL) Sensor	2020	04	12

Type of calibration	Date of calibration	H2S (ppm)	CO (ppm)	O2 (%)	LEL (%)
Sensor Calibration	2018 04 10	25	100	20.9	50
Result of Calibration		OK	OK	OK	OK

Calibration Cost: (As per attached invoice) F.O.C

Calibration remarks:

Next calibration date of this instrument will be : 2019 04 10

IMPORTANT NOTES TO BW GAS DETECTOR USERS

USERS MUST READ THE OPERATOR'S MANUAL THOROUGHLY BEFORE OPERATING THIS EQUIPMENT AND FOLLOW THEIR OWN SAFETY SUPERVISOR'S INSTRUCTION TO WORK.

All gas detection instrumentation on the market requires periodic calibration to accurately measure gas. Calibration is only as accurate as the test gas used. BW Technologies quality test gases are made to the highest accuracy and trace-ability to N. I.S.T. Standards.



Calibrated By: Sara Tse

Service Hotline: 2592 2120 Ms. Tse - Service Dept.

Asia Pacific Industrial Safety Equipment

Unit B, 1/F., Hing Yip Centre,
31 Hing Yip Street, Kwun Tong, Kowloon.
Tel: 2592 2100 Fax: 3165 8960

Asia Technologies

亞洲科技